

University of Washington

Standard Operating Procedures for Chemicals or Processes		
#1 Process (if applicable)	pH buffer 10 is primarily used for the calibration of the YSI EXO2 Multiparameter Water Quality Sonde (or “exosonde”)	
#2 Chemicals and Hazards	pH buffer 10 (colorless) Not hazardous.	
#3 Personal Protective Equipment (PPE)	Wear eye and face protection. Wear appropriate gloves and clothing to prevent skin exposure.	
#4 Environmental/Ventilation Controls	Should not be released into the environment (do not dispose down the drain.) Ensure adequate ventilation.	
#5 Special Handling Procedures & Storage Requirements	Wear PPE. Keep containers tightly closed in cool, dry, and well-ventilated environment.	
#6 Spill and Accident Procedures	Soak up with inert absorbent material. If eye contact, rinse immediately with plenty of water, also under eyelids, for at least 15 minutes. Eye wash station is located in center of wet lab, across from empty gas tanks. If skin contact, wash immediately with plenty of water for at least 15 minutes. Obtain medical attention if symptoms occur. If inhaled, move to fresh air immediately. If ingested, clean mouth out with water.	
#7 Waste Disposal	Dispose of contents in non-halogenated waste container, located in wet lab fume hood.	
#8 Special Precautions for Animal Use (if applicable)	*	
Particularly hazardous substance involved?	<input type="checkbox"/> YES:	Blocks #9 to #11 are Mandatory
	<input checked="" type="checkbox"/> NO:	Blocks #9 to #11 are Optional.
#9 Approval Required	Users must receive specific physical and health hazard information and safe laboratory work practices training from their supervisor.	
#10 Decontamination	Wash face, hands, and any exposed skin thoroughly after handling. If contact with skin, wash off immediately with plenty of water for at least 15 minutes. If inhaled, move to fresh air and, if symptoms persist, obtain medical attention.	
#11 Designated Area	HEEL wet lab, room 232.	
Name: _____ Title: _____		
Signature: _____ Date: _____		